


MODEL 3000 TEL-A-TENSION  
INDICATOR CALIBRATION TEST



3 WEIGHTS ARE USED IN  
THIS CALIBRATION TEST

50 GRAMS



100 GRAMS



400 GRAMS



TEST PROCEDURE

THE FOLLOWING IS A STATIC INDICATOR TEST  
THAT MAY BE USED TO CHECK CALIBRATION.

USING A THIN MONO-FILAMENT LINE,  
THREAD THE DEVICE IN THE NORMAL FASHION  
AS INDICATED IN THE ABOVE DRAWING. THEN  
FASTEN THE END OF LINE BY THE PRE-TENSION  
SECURELY AND ATTACH A SELECTED WEIGHT  
TO THE OTHER END OF THE LINE EXTENDING FROM  
THE TAKE-OFF WHEEL. NOW LIFT THE WEIGHT  
SLIGHTLY AND GENTLY LET IT DROP TO OVER  
COME THE FRICTION OF THE BALL BEARINGS.  
OBSERVE THE READING OF THE INDICATOR AND  
COMPARE IT WITH THE SELECTED WEIGHT. THE  
READINGS SHOULD COMPARE WITHIN PLUS OR  
MINUS 8 PERCENT

4-25-91

AZONIC PRODUCTS INC.

